

DIRECTIONS FOR USE**Edge K-Files Hand Instruments****Composition**

Edge K-Files™ are made with Stainless Steel.

Indications for Use

- Edge K-Files™ are used to form the initial GlidePath before EdgeEndo® NiTi rotary or reciprocating files.
- These instruments are for Dentist use only.

Contraindications

None known.

Warnings

- Use of a rubber dam system before use is advised.
- Edge K-Files™ are intended for single use only to avoid file separation.

Precautions

- Exercise caution in canals that merge, divide, and/or have curvatures.
- For personal safety, wear protective equipment such as gloves, glasses, mask, etc.
- Irrigate the root canal copiously and frequently during the procedure.
- Lubricants such as EdgeLube™ and EdgeGel™ are recommended while using all hand instruments such as K-Files.
- To establish the working length and initial GlidePath take a #10 hand file to the working length.

Precautions (cont.)

- Clean flutes frequently and check for signs of distortion or wear.
- Edge K-Files™ should be used in an in-out, watch-winding, or balanced-force motion.
- Edge K-Files™ are used to form the initial GlidePath and to follow the canal until the working length is achieved.

Adverse Reactions

None known.

Step by step instructions

- Form straightline access to the canal orifice.
- Use a #10 K-File to establish the working length with an apex locator and/or radiograph.
- Confirm patency with the #10 K-File.
- Irrigate frequently with EdgeLube, 17% Liquid EDTA.
- If the #10 K-File will not reach the working length, drop down and use a #6 K-File to the working length, then #8, and then #10 K-File.
- Irrigate with EdgeLube™, 17% Liquid EDTA between each file.
- After the #10 is at the working length and patency is confirmed either 1) Complete the GlidePath with GlidePath systems EdgeGlidePath™ or EdgeFind™ or 2) Start shaping the canal with EdgeEndo® rotary or reciprocating files.

**EDGEK-FILE™**